

Appendix F Water Quality Equipment Calibration Certificate



REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION

| | | | |
|----------|--|----------------|-------------|
| CONTACT: | BEN TAM | WORK ORDER: | HK1912056 |
| CLIENT: | ACTION UNITED ENVIRONMENT SERVICES AND CONSULTING | | |
| ADDRESS: | RM A 20/F., GOLD KING IND BLDG, NO. 35-41 TAI LIN PAI ROAD, KWAI CHUNG, N.T. HONG KONG | SUB-BATCH: | 0 |
| | | LABORATORY: | HONG KONG |
| | | DATE RECEIVED: | 20-Mar-2019 |
| | | DATE OF ISSUE: | 26-Mar-2019 |

COMMENTS

The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.

The "Tolerance Limit" quoted is the acceptance criteria applicable for similar equipment used by the ALS Hong Kong laboratory or quoted from relevant international standards.

| | |
|----------------------|---|
| Scope of Test: | Conductivity, Dissolved Oxygen, pH Value, Turbidity, Salinity and Temperature |
| Equipment Type: | Multifunctional Meter |
| Brand Name: | YSI |
| Model No.: | Professional DSS |
| Serial No.: | 17B102764/17B100758 |
| Equipment No.: | EQW019 |
| Date of Calibration: | 22 March, 2019 |

NOTES

This is the Final Report and supersedes any preliminary report with this batch number.

Results apply to sample(s) as submitted. All pages of this report have been checked and approved for release.

Ms. Lin Wai Yu
Assistant Manager - Inorganic

This report may not be reproduced except with prior written approval from ALS Technichem (HK) Pty Ltd.

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK1912056
SUB-BATCH: 0
DATE OF ISSUE: 26-Mar-2019
CLIENT: ACTION UNITED ENVIRONMENT SERVICES AND CONSULTING

Equipment Type: Multifunctional Meter
Brand Name: YSI
Model No.: Professional DSS
Serial No.: 17B102764/17B100758
Equipment No.: EQW019
Date of Calibration: 22 March, 2019 **Date of Next Calibration:** 22 June, 2019

PARAMETERS:

Conductivity Method Ref: APHA (21st edition), 2510B

| Expected Reading (µS/cm) | Displayed Reading (µS/cm) | Tolerance (%) |
|--------------------------|---------------------------|---------------|
| 146.9 | 143.1 | -2.6 |
| 6667 | 6194 | -7.1 |
| 12890 | 12016 | -6.8 |
| 58670 | 54263 | -7.5 |
| Tolerance Limit (%) | | ±10.0 |

Dissolved Oxygen Method Ref: APHA (21st edition), 4500-O: G

| Expected Reading (mg/L) | Displayed Reading (mg/L) | Tolerance (mg/L) |
|-------------------------|--------------------------|------------------|
| 4.81 | 4.63 | -0.18 |
| 6.77 | 6.60 | -0.17 |
| 8.33 | 8.28 | -0.05 |
| Tolerance Limit (mg/L) | | ±0.20 |

pH Value Method Ref: APHA (21st edition), 4500H:B

| Expected Reading (pH unit) | Displayed Reading (pH unit) | Tolerance (pH unit) |
|----------------------------|-----------------------------|---------------------|
| 4.0 | 4.07 | +0.07 |
| 7.0 | 7.19 | +0.19 |
| 10.0 | 10.04 | +0.04 |
| Tolerance Limit (pH unit) | | ±0.20 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Lin Wai Yu
 Assistant Manager - Inorganic

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK1912056
SUB-BATCH: 0
DATE OF ISSUE: 26-Mar-2019
CLIENT: ACTION UNITED ENVIRONMENT SERVICES AND CONSULTING

Equipment Type: Multifunctional Meter
Brand Name: YSI
Model No.: Professional DSS
Serial No.: 17B102764/17B100758
Equipment No.: EQW019
Date of Calibration: 22 March, 2019 **Date of Next Calibration:** 22 June, 2019

PARAMETERS:

Turbidity Method Ref: APHA (21st edition), 2130B

| Expected Reading (NTU) | Displayed Reading (NTU) | Tolerance (%) |
|------------------------|----------------------------|---------------|
| 0 | -0.24 | -- |
| 4 | 4.26 | +6.5 |
| 40 | 41.30 | +3.2 |
| 80 | 75.41 | -5.7 |
| 400 | 388.10 | -3.0 |
| 800 | 724.34 | -9.5 |
| | Tolerance Limit (%) | ±10.0 |

Salinity Method Ref: APHA (21st edition), 2520B

| Expected Reading (ppt) | Displayed Reading (ppt) | Tolerance (%) |
|------------------------|----------------------------|---------------|
| 0 | 0.00 | -- |
| 10 | 10.01 | +0.1 |
| 20 | 19.14 | -4.3 |
| 30 | 28.15 | -6.2 |
| | Tolerance Limit (%) | ±10.0 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Lin Wai Yu
 Assistant Manager - Inorganic

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK1912056
SUB-BATCH: 0
DATE OF ISSUE: 26-Mar-2019
CLIENT: ACTION UNITED ENVIRONMENT SERVICES AND CONSULTING

Equipment Type: Multifunctional Meter
Brand Name: YSI
Model No.: Professional DSS
Serial No.: 17B102764/17B100758
Equipment No.: EQW019
Date of Calibration: 22 March, 2019 Date of Next Calibration: 22 June, 2019

PARAMETERS:
Temperature Method Ref: Section 6 of International Accreditation New Zealand Technical Guide No. 3 Second edition March 2008: Working Thermometer Calibration Procedure.

| Expected Reading (°C) | Displayed Reading (°C) | Tolerance (°C) |
|-----------------------|------------------------|----------------|
| 8.5 | 10.0 | +1.5 |
| 23.0 | 22.4 | -0.6 |
| 41.0 | 39.1 | -1.9 |
| | Tolerance Limit (°C) | ±2.0 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Lin Wai Yu
Assistant Manager - Inorganic



REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION

| | | | |
|----------|--|----------------|-------------|
| CONTACT: | MR. POLAR CHAN | WORK ORDER: | HK1907349 |
| CLIENT: | ACUITY SUSTAINABILITY CONSULTING LIMITED | | |
| ADDRESS: | 1908, IPLACE, NOS. 301-305 CASTLE PEAK ROAD, Kwai Chung, New Territories, HONG KONG | SUB-BATCH: | 0 |
| | | LABORATORY: | HONG KONG |
| | | DATE RECEIVED: | 19-Feb-2019 |
| | | DATE OF ISSUE: | 28-Feb-2019 |

COMMENTS

The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.

The "Tolerance Limit" quoted is the acceptance criteria applicable for similar equipment used by the ALS Hong Kong laboratory or quoted from relevant international standards.

| | |
|----------------------|--|
| Scope of Test: | Dissolved Oxygen, pH Value, Turbidity, Salinity, Redox Potential and Temperature |
| Equipment Type: | Multifunctional Meter |
| Brand Name: | HORIBA |
| Model No.: | U-5000 |
| Serial No.: | UHB5F2BB |
| Equipment No.: | -- |
| Date of Calibration: | 26 February, 2019 |

NOTES

This is the Final Report and supersedes any preliminary report with this batch number.

Results apply to sample(s) as submitted. All pages of this report have been checked and approved for release.

Mr Chan Siu Ming, Vico
Manager - Inorganic

This report may not be reproduced except with prior written approval from ALS Technichem (HK) Pty Ltd.

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK1907349
 SUB-BATCH: 0
 DATE OF ISSUE: 28-Feb-2019
 CLIENT: ACUITY SUSTAINABILITY CONSULTING LIMITED

Equipment Type: Multifunctional Meter
 Brand Name: HORIBA
 Model No.: U-5000
 Serial No.: UHB5F2BB
 Equipment No.: --
 Date of Calibration: 26 February, 2019 Date of Next Calibration: 26 May, 2019

PARAMETERS:
 Dissolved Oxygen Method Ref: APHA (21st edition), 4500-O: G

| Expected Reading (mg/L) | Displayed Reading (mg/L) | Tolerance (mg/L) |
|-------------------------|--------------------------|------------------|
| 3.72 | 3.90 | +0.18 |
| 5.61 | 5.76 | +0.15 |
| 8.52 | 8.43 | -0.09 |
| | Tolerance Limit (mg/L) | ±0.20 |

pH Value Method Ref: APHA (21st edition), 4500H:B

| Expected Reading (pH unit) | Displayed Reading (pH unit) | Tolerance (pH unit) |
|----------------------------|-----------------------------|---------------------|
| 4.0 | 4.08 | +0.08 |
| 7.0 | 7.00 | +0.00 |
| 10.0 | 9.98 | -0.02 |
| | Tolerance Limit (pH unit) | ±0.20 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Mr Chan Siu Ming, Vico
 Manager - Inorganic

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK1907349
 SUB-BATCH: 0
 DATE OF ISSUE: 28-Feb-2019
 CLIENT: ACUITY SUSTAINABILITY CONSULTING LIMITED

Equipment Type: Multifunctional Meter
 Brand Name: HORIBA
 Model No.: U-5000
 Serial No.: UHB5F2BB
 Equipment No.: --
 Date of Calibration: 26 February, 2019 Date of Next Calibration: 26 May, 2019

PARAMETERS:

Turbidity Method Ref: APHA (21st edition), 2130B

| Expected Reading (NTU) | Displayed Reading (NTU) | Tolerance (%) |
|------------------------|-------------------------|---------------|
| 0 | 0.00 | -- |
| 4 | 4.17 | +4.3 |
| 40 | 39.8 | -0.5 |
| 80 | 78.4 | -2.0 |
| 400 | 398 | -0.5 |
| 800 | 784 | -2.0 |
| | Tolerance Limit (%) | ±10.0 |

Salinity Method Ref: APHA (21st edition), 2520B

| Expected Reading (ppt) | Displayed Reading (ppt) | Tolerance (%) |
|------------------------|-------------------------|---------------|
| 0 | 0.02 | -- |
| 10 | 9.37 | -6.3 |
| 20 | 18.06 | -9.7 |
| 30 | 27.41 | -8.6 |
| | Tolerance Limit (%) | ±10.0 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Mr Chan Siu Ming, Vico
Manager - Inorganic

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK1907349
 SUB-BATCH: 0
 DATE OF ISSUE: 28-Feb-2019
 CLIENT: ACUITY SUSTAINABILITY CONSULTING LIMITED

Equipment Type: Multifunctional Meter
 Brand Name: HORIBA
 Model No.: U-5000
 Serial No.: UHB5F2BB
 Equipment No.: --
 Date of Calibration: 26 February, 2019 Date of Next Calibration: 26 May, 2019

PARAMETERS:

Redox Potential Method Ref: APHA (21st edition), 2580B
 Method Ref: Orion Research Instruction Manual and the Laboratory Manual
 the Environmental of Water, Wastewater and Soil (2nd edition), Rump & Krist (1992)

| Expected Reading (mV) | Displayed Reading (mV) | Difference of A and B (mV) |
|-----------------------|------------------------|----------------------------|
| Solution A (~234mV) | 228 | |
| Solution B (~300mV) | 295 | +67.0 |
| | Tolerance Limit (mV) | >66 |

Temperature

Method Ref: Section 6 of International Accreditation New Zealand Technical
 Guide No. 3 Second edition March 2008: Working Thermometer Calibration Procedure.

| Expected Reading (°C) | Displayed Reading (°C) | Tolerance (°C) |
|-----------------------|------------------------|----------------|
| 11.0 | 11.48 | +0.5 |
| 21.0 | 21.66 | +0.7 |
| 38.0 | 39.07 | +1.1 |
| | Tolerance Limit (°C) | ±2.0 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Mr Chan Siu Ming, Vico
 Manager - Inorganic