

Appendix F Water Quality Equipment Calibration Certificate



REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION

| | | | |
|----------|--|----------------|-------------|
| CONTACT: | MR NELSON TSUI | WORK ORDER: | HK1917914 |
| CLIENT: | ACUITY SUSTAINABILITY CONSULTING LIMITED | | |
| ADDRESS: | 1908, iPLACE, NOS. 301-305 CASTLE PEAK ROAD, KWAI CHUNG, NEW TERRITORIES, HONG KONG | SUB-BATCH: | 0 |
| | | LABORATORY: | HONG KONG |
| | | DATE RECEIVED: | 30-Apr-2019 |
| | | DATE OF ISSUE: | 09-May-2019 |

COMMENTS

The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.

The "Tolerance Limit" quoted is the acceptance criteria applicable for similar equipment used by the ALS Hong Kong laboratory or quoted from relevant international standards.

The "Next Calibration Date" is recommended according to best practice principle as practised by the ALS Hong Kong laboratory or quoted from relevant international standards.

| | |
|----------------------|--|
| Scope of Test: | Dissolved Oxygen, pH Value, Turbidity, Salinity, Redox Potential and Temperature |
| Equipment Type: | Multifunctional Meter |
| Brand Name: | HORIBA |
| Model No.: | U-5000 Multiparameter Water Quality Meter |
| Serial No.: | -- |
| Equipment No.: | BGYP9CKD |
| Date of Calibration: | 09-May-2019 |

NOTES

This is the Final Report and supersedes any preliminary report with this batch number.

Results apply to sample(s) as submitted. All pages of this report have been checked and approved for release.

Mr Chan Siu Ming, Vico
Manager - Inorganic

This report may not be reproduced except with prior written approval from ALS Technichem (HK) Pty Ltd.

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK1917914
 SUB-BATCH: 0
 DATE OF ISSUE: 09-May-2019
 CLIENT: ACUITY SUSTAINABILITY CONSULTING LIMITED

Equipment Type: Multifunctional Meter
 Brand Name: HORIBA
 Model No.: U-5000 Multiparameter Water Quality Meter
 Serial No.: --
 Equipment No.: BGY99CKD
 Date of Calibration: 09-May-2019 Date of Next Calibration: 09-Aug-2019

PARAMETERS:
 Dissolved Oxygen Method Ref: APHA (21st edition), 4500-O: G

| Expected Reading (mg/L) | Displayed Reading (mg/L) | Tolerance (mg/L) |
|-------------------------|--------------------------|------------------|
| 8.10 | 8.11 | +0.01 |
| 5.57 | 5.42 | -0.15 |
| 3.15 | 3.06 | -0.09 |
| | Tolerance Limit (mg/L) | ±0.20 |

pH Value Method Ref: APHA (21st edition), 4500H:B

| Expected Reading (pH unit) | Displayed Reading (pH unit) | Tolerance (pH unit) |
|----------------------------|-----------------------------|---------------------|
| 4.0 | 3.94 | -0.06 |
| 7.0 | 7.06 | +0.06 |
| 10.0 | 9.87 | -0.13 |
| | Tolerance Limit (pH unit) | ±0.20 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Mr Chan Siu Ming, Vico
 Manager - Inorganic

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK1917914
 SUB-BATCH: 0
 DATE OF ISSUE: 09-May-2019
 CLIENT: ACUITY SUSTAINABILITY CONSULTING LIMITED

Equipment Type: Multifunctional Meter
 Brand Name: HORIBA
 Model No.: U-5000 Multiparameter Water Quality Meter
 Serial No.: --
 Equipment No.: BGYP9CKD
 Date of Calibration: 09-May-2019 Date of Next Calibration: 09-Aug-2019

PARAMETERS:

Turbidity Method Ref: APHA (21st edition), 2130B

| Expected Reading (NTU) | Displayed Reading (NTU) | Tolerance (%) |
|------------------------|-------------------------|---------------|
| 0 | 0.33 | -- |
| 4 | 4.21 | +5.3 |
| 40 | 42.3 | +5.7 |
| 80 | 81.3 | +1.6 |
| 400 | 402 | +0.5 |
| 800 | 845 | +5.6 |
| | Tolerance Limit (%) | ±10.0 |

Salinity Method Ref: APHA (21st edition), 2520B

| Expected Reading (ppt) | Displayed Reading (ppt) | Tolerance (%) |
|------------------------|-------------------------|---------------|
| 0 | 0.0 | -- |
| 10 | 9.8 | -2.0 |
| 20 | 19.7 | -1.5 |
| 30 | 29.9 | -0.3 |
| | Tolerance Limit (%) | ±10.0 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Mr Chan Siu Ming, Vico
Manager - Inorganic

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK1917914
SUB-BATCH: 0
DATE OF ISSUE: 09-May-2019
CLIENT: ACUITY SUSTAINABILITY CONSULTING LIMITED

Equipment Type: Multifunctional Meter
Brand Name: HORIBA
Model No.: U-5000 Multiparameter Water Quality Meter
Serial No.: --
Equipment No.: BGY99CKD
Date of Calibration: 09-May-2019 Date of Next Calibration: 09-Aug-2019

PARAMETERS:
Temperature Method Ref: Section 6 of International Accreditation New Zealand Technical Guide No. 3 Second edition March 2008: Working Thermometer Calibration Procedure.

| Expected Reading (°C) | Displayed Reading (°C) | Tolerance (°C) |
|-----------------------|------------------------|----------------|
| 9.5 | 10.11 | +0.6 |
| 19.0 | 19.45 | +0.4 |
| 38.5 | 39.05 | +0.5 |
| | Tolerance Limit (°C) | ±2.0 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Mr Chan Siu Ming, Vico
Manager - Inorganic

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK1917914
SUB-BATCH: 0
DATE OF ISSUE: 09-May-2019
CLIENT: ACUITY SUSTAINABILITY CONSULTING LIMITED

Equipment Type: Multifunctional Meter
Brand Name: HORIBA
Model No.: U-5000 Multiparameter Water Quality Meter
Serial No.: --
Equipment No.: BGY99CKD
Date of Calibration: 09-May-2019 Date of Next Calibration: 09-Aug-2019

PARAMETERS:
Redox Potential Method Ref: APHA (21st edition), 2580B
Method Ref: Orion Research Instruction Manual and the Laboratory Manual
the Environmental of Water, Wastewater and Soil (2nd edition), Rump & Krist (1992)

| Expected Reading (mV) | Displayed Reading (mV) | Difference of A and B (mV) |
|-----------------------|------------------------|----------------------------|
| Solution A (-234mV) | 68 | |
| Solution B (-300mV) | 115 | +47.0 |
| | Tolerance Limit (mV) | >66 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Mr Chan Siu Ming, Vico
Manager - Inorganic



專業化驗有限公司
QUALITY PRO TEST-CONSULT LIMITED

Unit 10, 14/F, Wah Wai Centre, 38-40 Au Pui Wan St., Fotan, Hong Kong
Email: info@qualityprotest.com; Website: www.qualityprotest.com
Tel: (852) 3956 8717; Fax: (852) 3956 3928

REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION

Report No. : AI070052
Date of Issue : 12 July, 2019
Page No. : 1 of 2

PART A – CUSTOMER INFORMATION

Acuity Sustainability Consulting Limited
Unit 1908, Nos. 301-305
Castle Peak Road, Kwai Chung
N.T., HK
Attn: Mr. Nelson TSUI

PART B – DESCRIPTION

Name of Equipment : YSI ProDSS (Multi-Parameters)
Manufacturer : YSI (a xylem brand)
Serial Number : 15M101091
Date of Received : Jul 09, 2019
Date of Calibration : Jul 11, 2019
Date of Next Calibration^(a) : Oct 11, 2019

PART C – REFERENCE METHODS/ DOCUMENTS FOR THE CALIBRATION

| Parameter | Reference Method |
|------------------|--|
| pH at 25°C | APHA 21e 4500-H ⁺ B |
| Dissolved Oxygen | APHA 21e 4500-O G |
| Salinity | APHA 21e 2520 B |
| Turbidity | APHA 21e 2130 B |
| Temperature | Section 6 of international Accreditation New Zealand Technical Guide no. 3 Second edition March 2008: Working Thermometer Calibration Procedure. |

PART D – CALIBRATION RESULTS^(b,c)

(1) pH at 25°C

| Target (pH unit) | Displayed Reading ^(d) (pH Unit) | Tolerance ^(e) (pH Unit) | Results |
|------------------|--|------------------------------------|--------------|
| 4.00 | 4.10 | 0.1 | Satisfactory |
| 7.42 | 7.28 | -0.14 | Satisfactory |
| 10.01 | 10.10 | 0.09 | Satisfactory |

Tolerance of pH should be less than ± 0.20 (pH unit)

(2) Temperature

| Reading of Ref. thermometer (°C) | Displayed Reading (°C) | Tolerance (°C) | Results |
|----------------------------------|------------------------|----------------|--------------|
| 15.0 | 15.0 | 0.0 | Satisfactory |
| 25.0 | 25.5 | 0.5 | Satisfactory |
| 52.0 | 53.7 | 1.7 | Satisfactory |

Tolerance limit of temperature should be less than ± 2.0 (°C)

~ CONTINUED ON NEXT PAGE ~

Remark(s): -

- ^(a) The "Date of Next Calibration" is recommended according to best practice principals as practiced by QPT or quoted from relevant international standards.
^(b) The results relate only to the calibrated equipment as received
^(c) The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.
^(d) "Displayed Reading" denotes the figure shown on item under calibration/ checking regardless of equipment precision or significant figures.
^(e) The "Tolerance Limit" mentioned is referenced to YSI product specifications.


LEE Chun-ning, Desmond
Senior Chemist



專業化驗有限公司

QUALITY PRO TEST-CONSULT LIMITED

Unit 10, 14/F, Wah Wai Centre, 38-40 Au Pui Wan St., Fotan, Hong Kong

Email: info@qualityprotest.com; Website: www.qualityprotest.com

Tel: (852) 3956 8717; Fax: (852) 3956 3928

REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION

Report No. : AI070052
Date of Issue : 12 July, 2019
Page No. : 2 of 2

PART D – CALIBRATION RESULTS (Cont'd)

(3) Dissolved Oxygen

| Expected Reading (mg/L) | Displayed Reading (mg/L) | Tolerance (mg/L) | Results |
|-------------------------|--------------------------|------------------|--------------|
| 7.68 | 7.75 | 0.07 | Satisfactory |
| 5.81 | 5.71 | -0.1 | Satisfactory |
| 3.20 | 3.38 | 0.18 | Satisfactory |
| 0.20 | 0.1 | -0.10 | Satisfactory |

Tolerance limit of dissolved oxygen should be less than ± 0.50 (mg/L)

(4) Salinity

| Expected Reading (g/L) | Displayed Reading (g/L) | Tolerance (%) | Results |
|------------------------|-------------------------|---------------|--------------|
| 10 | 9.96 | -0.4 | Satisfactory |
| 20 | 20.26 | 1.3 | Satisfactory |
| 30 | 30.97 | 3.2 | Satisfactory |

Tolerance limit of salinity should be less than ± 10.0 (%)

(5) Turbidity

| Expected Reading (NTU) | Displayed Reading ^(f) (NTU) | Tolerance ^(g) (%) | Results |
|------------------------|--|------------------------------|--------------|
| 0 | 0.24 | -- | Satisfactory |
| 10 | 10.58 | 5.8 | Satisfactory |
| 20 | 20.00 | 0.0 | Satisfactory |
| 100 | 97.60 | -2.4 | Satisfactory |
| 800 | 770.00 | -3.8 | Satisfactory |

Tolerance limit of turbidity should be less than ± 10.0 (%)

~ END OF REPORT ~

Remark(s): -

^(f) "Displayed Reading" presents the figures shown on item under calibration/ checking regardless of equipment precision or significant figures.

^(g) The "Tolerance Limit" mentioned is the acceptance criteria applicable for similar equipment used by Quality Pro Test-Consult Ltd. or quoted from relevant international standards.



REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION

| | | | |
|----------|--|----------------|-------------|
| CONTACT: | MR NELSON TSUI | WORK ORDER: | HK1923178 |
| CLIENT: | ACUITY SUSTAINABILITY CONSULTING LIMITED | | |
| ADDRESS: | 1908, iPLACE, NOS. 301-305 CASTLE PEAK ROAD, KWAI CHUNG, NEW TERRITORIES, HONG KONG | SUB-BATCH: | 0 |
| | | LABORATORY: | HONG KONG |
| | | DATE RECEIVED: | 05-Jun-2019 |
| | | DATE OF ISSUE: | 21-Jun-2019 |

COMMENTS

The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.

The "Tolerance Limit" quoted is the acceptance criteria applicable for similar equipment used by the ALS Hong Kong laboratory or quoted from relevant international standards.

The "Next Calibration Date" is recommended according to best practice principle as practised by the ALS Hong Kong laboratory or quoted from relevant international standards.

| | |
|----------------------|--|
| Scope of Test: | Dissolved Oxygen, pH Value, Turbidity, Salinity, Redox Potential and Temperature |
| Equipment Type: | Multifunctional Meter |
| Brand Name: | HORIBA |
| Model No.: | U-53 |
| Serial No.: | UHB5F2BB |
| Equipment No.: | -- |
| Date of Calibration: | 14-Jun-2019 |

NOTES

This is the Final Report and supersedes any preliminary report with this batch number.

Results apply to sample(s) as submitted. All pages of this report have been checked and approved for release.

Mr Chan Siu Ming, Vico
Manager - Inorganic

This report may not be reproduced except with prior written approval from ALS Technichem (HK) Pty Ltd.

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK1923178
 SUB-BATCH: 0
 DATE OF ISSUE: 21-Jun-2019
 CLIENT: ACUITY SUSTAINABILITY CONSULTING LIMITED

Equipment Type: Multifunctional Meter
 Brand Name: HORIBA
 Model No.: U-53
 Serial No.: UHB5F2BB
 Equipment No.: --
 Date of Calibration: 14-Jun-2019 Date of Next Calibration: 14-Sep-2019

PARAMETERS:
 Dissolved Oxygen Method Ref: APHA (21st edition), 4500-O: G

| Expected Reading (mg/L) | Displayed Reading (mg/L) | Tolerance (mg/L) |
|-------------------------|--------------------------|------------------|
| 2.46 | 2.63 | +0.17 |
| 5.46 | 5.51 | +0.05 |
| 7.56 | 7.57 | +0.01 |
| | Tolerance Limit (mg/L) | ±0.20 |

pH Value Method Ref: APHA (21st edition), 4500H:B

| Expected Reading (pH unit) | Displayed Reading (pH unit) | Tolerance (pH unit) |
|----------------------------|-----------------------------|---------------------|
| 4.0 | 3.96 | -0.04 |
| 7.0 | 7.06 | +0.06 |
| 10.0 | 10.06 | +0.06 |
| | Tolerance Limit (pH unit) | ±0.20 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Mr Chan Siu Ming, Vico
 Manager - Inorganic

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK1923178
 SUB-BATCH: 0
 DATE OF ISSUE: 21-Jun-2019
 CLIENT: ACUITY SUSTAINABILITY CONSULTING LIMITED

Equipment Type: Multifunctional Meter
 Brand Name: HORIBA
 Model No.: U-53
 Serial No.: UHB5F2BB
 Equipment No.: --
 Date of Calibration: 14-Jun-2019 Date of Next Calibration: 14-Sep-2019

PARAMETERS:

Salinity Method Ref: APHA (21st edition), 2520B

| Expected Reading (ppt) | Displayed Reading (ppt) | Tolerance (%) |
|------------------------|-------------------------|---------------|
| 0 | 0.00 | -- |
| 10 | 9.82 | -1.8 |
| 20 | 19.06 | -4.7 |
| 30 | 30.62 | +2.1 |
| | Tolerance Limit (%) | ±10.0 |

Redox Potential Method Ref: APHA (21st edition), 2580B

Method Ref: Orion Research Instruction Manual and the Laboratory Manual
 the Environmental of Water, Wastewater and Soil (2nd edition), Rump & Krist (1992)

| Expected Reading (mV) | Displayed Reading (mV) | Difference of A and B (mV) |
|-----------------------|------------------------|----------------------------|
| Solution A (-234mV) | 226 | |
| Solution B (-300mV) | 293 | +67.0 |
| | Tolerance Limit (mV) | >66 |

Temperature Method Ref: Section 6 of International Accreditation New Zealand Technical
 Guide No. 3 Second edition March 2008: Working Thermometer Calibration Procedure.

| Expected Reading (°C) | Displayed Reading (°C) | Tolerance (°C) |
|-----------------------|------------------------|----------------|
| 11.5 | 13.42 | +1.9 |
| 23.0 | 24.94 | +1.9 |
| 40.0 | 40.18 | +0.2 |
| | Tolerance Limit (°C) | ±2.0 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Mr Chan Siu Ming, Vico
 Manager - Inorganic

REPORT OF EQUIPMENT CALIBRATION



WORK ORDER: HK1923178
SUB-BATCH: 0
DATE OF ISSUE: 21-Jun-2019
CLIENT: ACUITY SUSTAINABILITY CONSULTING LIMITED

Equipment Type: Multifunctional Meter
Brand Name: HORIBA
Model No.: U-53
Serial No.: UHB5F2BB
Equipment No.: --
Date of Calibration: 20-Jun-2019 Date of Next Calibration: 14-Sep-2019

PARAMETERS:
Turbidity Method Ref: APHA (21st edition), 2130B

| Expected Reading (NTU) | Displayed Reading (NTU) | Tolerance (%) |
|------------------------|-------------------------|---------------|
| 0 | 0.00 | -- |
| 4 | 3.92 | -2.0 |
| 40 | 39.1 | -2.3 |
| 80 | 79.3 | -0.9 |
| 400 | 392 | -2.0 |
| 800 | 798 | -0.3 |
| | Tolerance Limit (%) | ±10.0 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Mr Chan Siu Ming, Vico
Manager - Inorganic

REPORT OF EQUIPMENT PERFORMANCE CHECK



WORK ORDER: HK1923178
SUB-BATCH: 0
DATE OF ISSUE: 21-Jun-2019
CLIENT: ACUITY SUSTAINABILITY CONSULTING LIMITED

Equipment Type: Multifunctional Meter
Brand Name: HORIBA
Model No.: U-53
Serial No.: UHB5F2BB
Equipment No.: --
Date of Calibration: 14-Jun-2019

PARAMETERS:

Turbidity Method Ref: APHA (21st edition), 2130B

| Expected Reading (NTU) | Displayed Reading (NTU) | Tolerance (%) |
|------------------------|-------------------------|---------------|
| 0 | 0.67 | -- |
| 4 | 23.8 | +495.0 |
| 40 | 150 | +275.0 |
| 80 | 222 | +177.5 |
| 400 | 729 | +82.3 |
| 800 | 0.00 | -100.0 |
| | Tolerance Limit (%) | ±10.0 |

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Mr Chan Siu Ming, Vico
Manager - Inorganic